



CONCORD MEDICAL LIMITED



www.concord-medical.com



00852-28666308



00852-35425676



86-0769-88919066



86-0769-88919050

EO Sterilization Report

QR No. : QR720-09

Date : 2021-4-27
Client : Mesoskinline ApS
Add : Juelsmindevej 57, DK-7120 Vejle,
Denmark, Zip code: DK-7120
Contact Person: Per Kjaersfeldt
Company NO.:

Ref No. CCeo-21042603
Tel No.:
Fax:
E-mail:

Goods / Product No.	QTY	Package	Size	Unit	Qty	Date	Production Lot No.
Plastic tips Model numbers: MST3001, MST3002, MST3003, MST3004	78,400 PCS	64 ctns	4.474 M ³	Batch	1	2021-4-27	2021-14-001 2021-14-002 2021-14-003 2021-14-004
Total package/volume		64 ctns	4.474 M ³	Batch	One	Toatal Lot No.	One

Full-Cycle sterilization parameters

Temp °C	Humidity	Vacuum	EO	Expo Min	Post Vacuum	Pressure	Air in bleed	Releasing
50~55	40~80%	-65±1 Kpa	600g/m ³	480	3	-30±1 Kpa	30 Min	7 day

Delivered Full-Cycle sterilization parameters

Temp °C	Humidity	Vacuum	EO	Expo Min	Post Vacuum	Pressure	Air in bleed	Releasing
50.1	64.8%	-65±1 Kpa	6 KG	480	3	-30±1 Kpa	30 Min	7 day

We hereby certify that the above mentioned goods (as described by the client) have been sterilized by "Ethylene Oxide" according to regulation of EN ISO 11135

Remark : Sterilization validation Yes ☐ None ☒

Operator: Lin Peihong

QC:

Approved by :

Date : 2021-4-27

Date: 2021-4-27

Date : 2021-4-27

ISO 13485 : 2016 and ISO 11135 Certified. E-mail Add : info@concord-medical.com

Factory HK: Workshop J, Shield Commercial Bldg, 84-92 Chai Wan Kok St, Tsuen Wan, Hong Kong

Factory Add : Floor 4, Block C, Yue Neng Ind Park, Fu Ming Road, Xia Bian Village, Houjie, Dong Guan, China. Zip 523960